Order Form - FlexNeo MAX

Prosthetic socket circumference at the very end of the cover in [cm]

any value every 0.5 cm; e.g. 35.5 cm; 37 cm ...



To take the measurement correctly, please mark the height of the cover in several places and place a ruler below the marking. The measurement should be made very carefully because moving the ruler even by a few millimetres up or down causes an incorrect reading of the circumference even by several centimetres.




The height of the cover in [cm]





The height of the cover — must take into account the gap between the cover and the foot. Usually, it is 7 mm. The softer the foot (e.g. fibreglass), the greater the gap will be.



Foot shell width in [cm]

Foot shell length in [cm]

Foot manufacturer, type and size

For example : OB Taleo 1C50 L29



The silicone sleeve - option

Cover below the silicone sleeve or without

Cover over the silicone sleeve

In this case, make sure you are taking the measurement of circuit with the sleeve on the socket.

Adapter			Colour		
30		SkinTone	Red		
34	П	Black	☐ White		

Skin Colour 1..29

Tattoo

Catalog no. for example: 08-1001 https://nextt.pl/pl/flex-neo/ak-design



BK - Transtibial



Photo of the foot from a distance of about 1 meter

camera lens position is very important



Four photos of the prosthesis on each side

front, back, right and left. The camera lens is approximately 30 cm above the floor.

There must not be any shoes, socks or other pieces of clothing included in the photo.









Scan of the prosthesis and the patient's leg

in a standing position, without the shoes, socks and other items of clothing - *.stl; *.obj

